

## Precision meets Beauty

# NanoPhotometer<sup>®</sup> N60 Microvolume Spectroscopy



**Microvolume Capability Built-in Vortex** Starting with only 0.3 μl of sample



**Full Scan** 3.5 seconds per reading 200 to 900 nm Resolution better than 1.8 nm



## Certainty in Real Time and IQ/OQ Package

Blank Control<sup>™</sup>, air bubble and impurity recognition Compliant with international standards in regulated environments





### **Endless Connectivity**

Built-in File Server for data access from Windows and Mac computers Print to Airprint<sup>™</sup> and HP Universal Driver compatible printers as well as DYMO Label printers Rest API for LIMS integration



Battery Powered Up to 10 hours battery operation



### Flexible Unit Control and Ultimate Data Security

Computer (Windows & Mac) Built-in touchscreen Smartphone / Tablet (Android OS & iOS) Proprietary NPOS immune to known threats

World's smallest footprint in its class: only 20 x 20 x 12 cm Ideal for nucleic acids, protein and samples in most organic solvents Allows kinetic studies in a drop No reconditioning, no recalibration and no regular maintenance ever Stand-alone operation with built-in 7 inch glove compatible touch screen Universal data output: Excel and PDF Multi Language User Interface Barcode ready 32 GB of onboard memory

### **Technical Specifications**

NanoVolume Performance		Zero Stability	±0.003 A/hour after 20 min warm up @ 280 nm
Detection Range dsDNA	1 ng/μl to 16,500 ng/μl (N50: 5 ng/μl to 7,500 ng/μl)	Noise	0.002 A rms at 0 A @ 280 nm 0.002 A (pk to pk) at 0 A @ 280 nm
Detection Range BSA	0.03 mg/ml to 478 mg/ml (N50: 0.15 mg/ml to 217 mg/ml)	Optical Arrangement	1 x 3648 CCD Array (N50: 1 x 1024 CCD Array)
Minimum Sample Size	0.3 µl	Lamp	Xenon flash lamp
Photometric Range (10 mm equivalent)	0.02 - 330 A (N50: 0.1 - 150 A)	Lifetime	$10^9$ flashes, up to 10 years
Path Length	0.67 and 0.07 mm	Processing Power & Compatibility	
Dilution Factor	15 and 140	Operating System	Linux based NPOS
Vortexer	2,800 rpm; tube size up to 2.0 ml	Onboard Processor	Quad Core 1 GHz
		Internal Storage	32 GB
Cuvette Performance Detection Range dsDNA	0.1 ng/μl to 130 ng/μl	Control Options	Onboard with built-in Touchscreen, Computer, Smartphone and Tablet
Detection Range BSA	0.003 mg/ml to 3.7 mg/ml	Software Compatibility	Windows 7, 8, 10 (32 & 64 bit), OS X, iOS & Android OS
Photometric Range	0 - 2.6 A	Min. Requirement Smartphone/Tablet	4" screen; Apple: iPad 2, iPhone5 & iOS 6; Android Phone: OS version 4.4; Android Tablet: OS version 5.0, Quadcore 1.2 GHz with 1 GB RAM
Center Height (Z-Height)	8.5 mm		
Cell Types	Outside dimension 12.5 x 12.5 mm	General Specifications	
Heating $37 \degree C \pm 0.5 \degree C$		Main Body Size	20 cm x 20 cm x 12 cm
Optical Specifications		Weight	3.8 - 5.2 kg depending on configuration
Wavelength Scan Range	200 - 900 nm (N50: 200 - 650 nm)		
Measure Time For Full Scan Range	3.5 - 6.0 seconds	Operating Voltage	90 - 250 V, 50/60 Hz, 60 W (90 W with battery pack), 18/19 VDC
Wavelength Reproducibility	± 0.2 nm (N50: ± 1 nm)	Display	1024 x 600 pixels; Touchscreen glove compatible
Wavelength Accuracy	± 0.75 nm (N50: 1.5 nm)	Built-in Battery Pack	Optional rechargeable lithium ion battery; 95 Wh, 6.6 Ah; Operation time: up to 10 h; min. charging cycles: 800
Bandwidth	better than 1.8 nm (N50: 5 nm)		
Stray Light	< 0.5 % at 240 nm using NaI (N50: < 2 %) and < 1 % at 280 nm using Acetone (N50: < 2 %)	Certification	CE, IEC 61010-1:2012 and EN 61326-1:2013
		Battery Certification	IEC 62133 and UN38.3 transport test
Absorbance Reproducibility	< 0.002 A (0.67 mm path) @ 280 nm (N50: < 0.004 A (0.67 mm path) @ 280 nm)	In & Output Ports	2x USB A, USB B, HDMI, Ethernet, WLAN
Absorbance Accuracy	< 1.75 % @ 0.7 A (0.67 mm path) @ 280 nm of the reading	Additional Data Input	Mouse & keyboard options
		Security	Slot for Kensington lock

### **Reviews**

"Awesome machine. I would purchase another one for additional labs."

Rating: 5.0  $\star \star \star \star \star$ 

Application Area: Genetics Academic Laboratory - Microarray Core

"I love the dynamic range for RNA/DNA measurements. We did our own in house check for reproducibility. The interface is very user friendly and easier to use than ... We like that we can use 1 ul of precious sample for an accurate reading rather than the required 1.5ul for ... (...) This has been a god-send. We have very low concentration samples that are very precious and this allows us to make measurements on these types of samples. Also, after doing PCR amplification, we no longer have to make dilutions for the upper limit readings due to the large dynamic range."

Twyla Juehne

Organization: Washington University School of Medicine

#### "Great machine with great results"

#### Rating: 5.0 $\star \star \star \star \star$

Application Area: Analysis of RNA, DNA, and protein concentrations

"This is an easy to use machine that gives great results. We have run it against several standard curves. Would definitely recommend it."

George Perry Organization: South Dakota State University