Technical Service Consultants Ltd

Microbiology House, Fir Street, Heywood, Lancashire OL10 1NW, UK T: +44 (0)1706 620600 F: +44 (0)1706 620445

E:sales@tscswabs.co.uk W:www.tscswabs.co.uk

Product Specification Sheet

TS/19-ASwab with a polystyrene shaft tipped with pure viscose, in polypropylene outer tube

Product Code	TS/19-A
Description	Swab with a polystyrene shaft tipped with pure viscose, in polypropylene outer tube
Intended Use	Culture swab; Patient sampling & general Laboratory use
Components	Polypropylene Tube Polyethylene cap viscose tip/polystyrene shaft swab (tip diameter – 4.5-5.0 mm) Non-toxic glue Label & film
Contents	Not applicable.
Method of Sterilisation	Sterilised by gamma irradiation at dosage of 25-46kGy
Shelf Life	24 months from date of manufacture
Storage Conditions	4-25°C. Store in dry and ventilated area; protect from light, heat and moisture in tightly closed original container
Precautions and Limitations	
Pack Size	250
Other	Quality control testing details on final product:i) Validation of irradiation sterility.

This specification is intended for guidance only.

You should always refer to a professional with experience in your intended application if in any doubt.

Issue #: 4 Date: 06.06.2016 ProductSpec_TS19-A